



HADLEIGH INNOVATIONS

Calibration report for Digital Thermometer, component of Plumtest TMV Test Kit

Products	Digital Thermometer DTM 305 B and TPK-03 S Probe
Brand	Plumtest
Test Conditions	20°C ± 2°C
Tolerance	± (0.3%+1°C) of reading for the calibration point at 60°C

The results of the tests, calibrations included in this document are traceable to National/Int'l Standards.

Calibration at 60.0°C:

Thermometer	Probe	Reading °C	Error °C
#1 s/n: 2503641	#1 TP25K0188	60.1	0.1
#2 s/n: 2503642	#2 TP25K0189	60.1	0.1
#3 s/n: 2503643	#3 TP25K0190	59.8	-0.2
#4 s/n: 2503644	#4 TP25K0192	60.1	0.1
#5 s/n: 2503645	#5 TP25K0198	60.0	0.0
#6 s/n: 2503646	#6 TP25K0206	60.1	0.1
#7 s/n: 2503647	#7 TP25K0207	60.1	0.1
#8 s/n: 2503648	#8 TP25K0214	60.0	0.0
#9 s/n: 2503649	#9 TP25K0215	60.0	0.0
#10 s/n: 2503650	#10 TP25K0216	60.1	0.1

General Guide AS ISO/IEC 17025:2017
NMI Monograph 5: Thermocouples in Temperature Measurement.

For and on behalf of

Hadleigh Innovations Pty Ltd

Authorised signature